

Fahrenheit Temperature Readings Only

CORRECT TEMPERATURE RANGE FOR REFRIGERATOR IS 35-46 DEGREES FAHRENHEIT

FAHRENHEIT TEMPERATURE LOG		Month: _____																												Days 1-31		
Refrigerator																																
Staff																																
Initials																																
Time																																
Day of the																																
Month	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24	25	26	27	28	29	30	31	
	am	pm	am	pm	am	pm	am	pm	am	pm	am	pm	am	pm	am	pm	am	pm	am	pm	am	pm	am	pm	am	pm	am	pm	am	pm	am	pm
Temp																																
48	Take Corrective Action if temperature is in this Red Area										Take Corrective Action if temperature is in this Red Area										Take Corrective Action if temperature is in this Red Area											
47	Take Corrective Action if temperature is in this Red Area										Take Corrective Action if temperature is in this Red Area										Take Corrective Action if temperature is in this Red Area											
46																																
45																																
44																																
43																																
42																																
41																																
40																																
39																																
38																																
37																																
36																																
35																																
34	Take Corrective Action if temperature is in this Red Area										Take Corrective Action if temperature is in this Red Area										Take Corrective Action if temperature is in this Red Area											
33	Take Corrective Action if temperature is in this Red Area										Take Corrective Action if temperature is in this Red Area										Take Corrective Action if temperature is in this Red Area											
CORRECT TEMPERATURE RANGE FOR FREEZER IS LESS THAN OR EQUAL TO 5 DEGREES FAHRENHEIT																																
FREEZER																																
7	Take Corrective Action if temperature is in this Red Area										Take Corrective Action if Temperature is in this Red Area										Take Corrective Action if temperature is in this Red Area											
6	Take Corrective Action if temperature is in this Red Area										Take Corrective Action if Temperature is in this Red Area										Take Corrective Action if temperature is in this Red Area											
5																																
4																																
3																																
2																																
1																																

Instructions: Place an "X" in the box that corresponds with the temperature (rows), day of the month, and am or pm (columns) for your temperature check. Then enter your initials and the time you monitored the temperature in the boxes at the top of the chart.

If the temperature is in the red area:

1. Store the vaccine under proper conditions as quickly as possible.
2. Isolate and label vaccine in the refrigerator/freezer. DO NOT USE until viability of vaccine has been determined by the manufacturers and Colorado Department of Public Health and Environment (CDPHE).
3. Call the vaccine manufacturers to determine whether the potency of the vaccines has been affected.
4. Document the date of the incident and what corrective action was taken.
5. Call the CDPHE Immunization Program for further assistance at (303) 692-2798 or (303) 692-2334.

Notes: _____

Vaccines Manufacturers:

- Adventis Pasteur, Inc. (800) 822-2463
- GlaxoSmithKline: (800) 366-8900
- Merck & CO.: (800) 672-6372
- Wyeth Lederle: (800) 358-7443